



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01644	Calib. Date: 04/21/2021
Certification File No.:	01645	Calib. No.: 00040
Linearity File No.:	01646	Cert. Date: 04/21/2021
Solution File No.:	01685	Cert. No.: 00033
Sequential File No.:	01685	Lin. Date: 04/21/2021
		Lin. No.: 00033
		Soln. Date: 07/31/2021
		Soln. No.: 00246
		File Date: 07/31/2021

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUD S3-0015
Control Solution %:	0.100%		Expires: 09/09/2021
Solution Control Lot:	19240		Bottle No.: 1074

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:50D		
Control 1 EC	0.098%	21:50D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:51D		
Control 2 EC	0.097%	21:52D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:52D		
Control 3 EC	0.098%	21:53D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUNA

First Name: JOSEPH

MI: M

Signature: _____

Badge No.: 112

Date: 07/31/2021