

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARSA-0031
Location: STAFFORD TOWNSHIP P.D.
Calibration File No.: 01838 Calib. Date: 06/14/2023 Calib. No.: 00045
Certification File No.: 01839 Cert. Date: 06/14/2023 Cert. No.: 00038
Linearity File No.: 01840 Lin. Date: 06/14/2023 Lin. No.: 00038
Solution File No.: 01855 Soln. Date: 07/07/2023 Soln. No.: 00275
Sequential File No.: 01855 File Date: 07/07/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUD S3-0014
Control Solution %: 0.100% Expires: 09/01/2023
Solution Control Lot: 21360 Bottle No.: 1401

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:57D		
Control 1 EC	0.097%	07:57D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:58D		
Control 2 EC	0.097%	07:58D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:59D		
Control 3 EC	0.097%	08:00D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: POVLOSKY

First Name: MARC

MI: R

Signature: _____



Badge No.: 111

Date: 07/07/2023