

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	Calib. No.:	00045
Calibration File No.:	01838	Calib. Date:	06/14/2023
Certification File No.:	01839	Cert. Date:	06/14/2023
Linearity File No.:	01840	Lin. Date:	06/14/2023
Solution File No.:	01866	Soln. Date:	08/18/2023
Sequential File No.:	01866	File Date:	08/18/2023

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUD S3-0014
Control Solution %:	0.100%			Expires:	01/11/2025
Solution Control Lot:	23010			Bottle No.:	0063

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:17D		
Control 1 EC	0.100%	23:17D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:18D		
Control 2 EC	0.098%	23:19D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:19D		
Control 3 EC	0.098%	23:20D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BUSH

First Name: JACKMAN

MI: R

Signature: \_\_\_\_\_



Badge No.: 127

Date: 08/18/2023