

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01838	Calib. Date: 06/14/2023
Certification File No.:	01839	Calib. No.: 00045
Linearity File No.:	01840	Cert. Date: 06/14/2023
Solution File No.:	01875	Cert. No.: 00038
Sequential File No.:	01875	Lin. Date: 06/14/2023
		Lin. No.: 00038
		Soln. Date: 09/11/2023
		Soln. No.: 00279
		File Date: 09/11/2023

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUD S3-0015
Control Solution %: 0.100%		Expires: 01/11/2025
Solution Control Lot: 23010		Bottle No.: 0395

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:08D		
Control 1 EC	0.099%	00:09D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:09D		
Control 2 EC	0.098%	00:10D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:11D		
Control 3 EC	0.098%	00:11D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RACANELLI

First Name: NICHOLAS

MI: M

Signature:



Badge No.: 122

Date: 09/11/2023