



Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01838	Calib. Date: 06/14/2023
Certification File No.:	01839	Cert. No.: 00045
Linearity File No.:	01840	Cert. Date: 06/14/2023
Solution File No.:	01867	Lin. No.: 00038
Sequential File No.:	01867	Soln. Date: 08/21/2023
		Soln. No.: 00278
		File Date: 08/21/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0015
Solution Control Lot:	23010	Expires: 01/11/2025
		Bottle No.: 0338

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:05D		
Control 1 EC	0.101%	12:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:06D		
Control 2 EC	0.100%	12:06D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:07D		
Control 3 EC	0.099%	12:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:08D		


All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKP2-304 DL

Changed By:

Last Name: LUTZ First Name: DENNIS MI: J

Signature:  Badge No.: 7045
 Date: 08/21/2023

DEPARTMENT OF
Justice and Public Safety
 Title to be certified in
Dennis J. Lutz
 Breath Test Coordinator/Instructor

THIS IS TO CERTIFY THAT THE ABOVE NAMED INDIVIDUAL HAS COMPLETED THE COURSE OF INSTRUCTION IN THE STATE OF NEW JERSEY IN THE COURSE OF BREATH TEST COORDINATOR/INSTRUCTOR AS REQUIRED BY THE BOARD OF PUBLIC SAFETY AND THE BOARD OF EXAMINERS OF THE STATE OF NEW JERSEY.

THE BOARD OF PUBLIC SAFETY HAS REVIEWED THE RECORDS OF THE INDIVIDUAL AND HAS DETERMINED THAT THE INDIVIDUAL IS QUALIFIED TO PERFORM THE DUTIES OF THE POSITION OF BREATH TEST COORDINATOR/INSTRUCTOR.

[Signature] SUPERVISOR
[Signature] ATTORNEY GENERAL

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1.		
2.		
3.		
4.		
5.		
6.		
7.		
8.		
9.		

K.P. 2030 (Rev. 01/98)

DEPARTMENT OF
Justice and Public Safety
 Title to be certified in
New Jersey State Police

THIS IS TO CERTIFY THAT THE ABOVE NAMED INDIVIDUAL HAS COMPLETED THE COURSE OF INSTRUCTION IN THE STATE OF NEW JERSEY IN THE COURSE OF BREATH TEST COORDINATOR/INSTRUCTOR AS REQUIRED BY THE BOARD OF PUBLIC SAFETY AND THE BOARD OF EXAMINERS OF THE STATE OF NEW JERSEY.

THE BOARD OF PUBLIC SAFETY HAS REVIEWED THE RECORDS OF THE INDIVIDUAL AND HAS DETERMINED THAT THE INDIVIDUAL IS QUALIFIED TO PERFORM THE DUTIES OF THE POSITION OF BREATH TEST COORDINATOR/INSTRUCTOR.

THE THOUSAND AND Nine DAY OF October

[Signature] SUPERVISOR
[Signature] ATTORNEY GENERAL

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1. 2-3-11	OCPA	Wm Hon
2. 1/24/13	OCPA	Adam Standa
3. 11-23-15	GCPA	M. Gonzalez
4. 4/6/17	LOVEWELL	Adam Standa
5. 9/22/19	NJSP Gallegos	Bob
6. 7-19-21	Lakehurst	Jim Kelly
7.		
8.		
9.		

K.P. 2030 (Rev. 07/07)

Dräger

Simulator

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)
Dräger, Inc.

- Model: ALCOTEST CU34
- Model: MARK IIA
- X-Cal 2000 (Alcosim)
- Other: _____

Serial Number:

DDUDS3-0015

Certification Date:
7-5-23

Technician:
DH / MB

Re-Certification Due Date:
7-5-24

Dräger

Alcotest 7110 Temperature Probe

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your state's specifications. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest 7110.

Serial Number Temp Probe:
DDXKP2-304

Certification Date:
7-5-23

Next Certification Due:
7-5-24

Probe Value:
106

Dräger, Inc. DH / MB



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

PHILIP D. MURPHY
Governor

SHEILA Y. OLIVER
Lt. Governor

MATTHEW J. PLATKIN
Attorney General

COLONEL PATRICK J. CALLAHAN
Superintendent

CERTIFICATION OF ANALYSIS
0.100 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger, Inc.

ANALYSIS DATE: 01/25/2023

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 23010

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1211 to 0.1225 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 11, 2025.

As OFS Director for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Michael Kennedy
Michael Kennedy
OFS Director
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 30 day of January, 2023.

Notary

KAREN E. STAHL
NOTARY PUBLIC OF NEW JERSEY
Commission # 60110522
My Commission Expires 8/13/2024



An Internationally Accredited Agency

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable

