

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01838	Calib. Date: 06/14/2023
Certification File No.:	01839	Cert. No.: 00045
Linearity File No.:	01840	Lin. Date: 06/14/2023
Solution File No.:	01886	Cert. No.: 00038
Sequential File No.:	01886	Lin. No.: 00038
		Soln. Date: 10/13/2023
		Soln. No.: 00281
		File Date: 10/13/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0015
Solution Control Lot:	23010	Expires: 01/11/2025
		Bottle No.: 0399

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:59D		
Control 1 EC	0.100%	22:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:01D		
Control 2 EC	0.098%	22:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:02D		
Control 3 EC	0.098%	22:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: JOHNSON

First Name: TANNER

MI: J

Signature: \_\_\_\_\_



#130

Badge No.: 130

Date: 10/13/2023