

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARSA-0031
Location:	STAFFORD TOWNSHIP P.D.	
Calibration File No.:	01838	Calib. Date: 06/14/2023
Certification File No.:	01839	Calib. No.: 00045
Linearity File No.:	01840	Cert. Date: 06/14/2023
Solution File No.:	01880	Cert. No.: 00038
Sequential File No.:	01880	Lin. Date: 06/14/2023
		Lin. No.: 00038
		Soln. Date: 10/01/2023
		Soln. No.: 00280
		File Date: 10/01/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUD S3-0015
Solution Control Lot:	23010	Expires: 01/11/2025
		Bottle No.: 0585

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:59D		
Control 1 EC	0.099%	08:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00D		
Control 2 EC	0.099%	09:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:01D		
Control 3 EC	0.099%	09:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SZWED

First Name: TYLER

MI: R

Signature: _____

Badge No.: 129

Date: 10/01/2023